

**To: Segi-AT; Korkowa 24A; 04-502 Warsaw**

### **AOX STANDARD TESTING**

Job ref.:

19-25923; 19-25920; 19-25916; 19-25426; 19-25421; 19-25416; 19-25399; 19-24857; 19-24854; 19-24849; 19-24846

AOX standard testing:

Nominal value on the bottle is 1.03 mg/L with a tolerance of 0.77- 1.29 mg/L.  
Standard ordered from Hach where the kits we use are from.

Date	AOX Results (mg/L)
31/10/19	0.97
31/10/19	0.96
31/10/19	1.14
31/10/19	1.24
04/10/19	1.26
04/10/19	0.97
04/11/19	1.06
04/11/19	1.07
05/11/19	0.87
05/11/19	0.89
05/11/19	0.91
05/11/19	1.06



Glynn Harvey  
Laboratory Manager  
20.12.2019