

Table of contents

Field Steam Sterilizers (W6369-20A005/A)

- **Podmiot:** Department of National Defence
- **Kraj:** Kanada - Ontario
- **Data zakończenia:** 2019-08-23

Field Steam Sterilizers (W6369-20A005/A)

ID	A-CA-20190808-P082628
Podmiot	Department of National Defence
Data rozpoczęcia	2019-08-08
Data zakończenia	2019-08-23
Kraj	Kanada - Ontario
Język	Angielski

Field Steam Sterilizers (W6369-20A005/A)

Reference number PW-\$PET-907-1590

Solicitation number W6369-20A005/A

Publication date 2019-08-08

Date closing 2019-08-23 14:00 Eastern Daylight Time (EDT)

Publishing status Active

GSIN "N6530SB - STERILIZER, N6530SBG - STERILIZER, WATER TYPE

Region of delivery Ontario

Notice type PAC-ACAN

Trade agreement "World Trade Organization-Agreement on Government Procurement (WTO-AGP), Canada-Panama Free Trade Agreement, Canada-Korea Free Trade Agreement (CKFTA), Canadian Free Trade Agreement (CFTA), Comprehensive Economic and Trade Agreement (CETA), Comprehensive and Progressive Agreement for Trans-Pacific Partnership (CPTPP), Canada-Colombia Free Trade Agreement, Canada-Peru Free Trade Agreement (CPFTA), North American Free Trade Agreement (NAFTA)

Tendering procedure Generally only one firm has been invited to bid

Competitive procurement strategy

Procurement entity Public Works and Government Services Canada

End user entity Department of National Defence

Contact Information "Lamorie, Cindy, cynthia.lamorie@pwgsc.gc.ca, (613) 687-6655 (), (613) 687-6656, Petawawa Procurement Building S-111, Rm C-114 101 Menin Rd. Garrison Petawawa Petawawa Ontario K8H 2X3 Telephone No.: (613) 687-6655 () Facsimile No.: (613) 687-6656, Petawawa Procurement Building S-111, Rm C-114 101 Menin Rd. Garrison Petawawa Petawawa Ontario K8H 2X3 N° de téléphone : (613) 687-6655 () N° de télécopieur : (613) 687-6656

Description

Trade Agreement: CETA / WTO-AGP / CPTPP / NAFTA / CFTA / FTAs with Peru / Colombia / Panama / Korea Tendering Procedures: Generally only one firm has been invited to bid Non-Competitive Procurement Strategy: Exclusive Rights Comprehensive Land Claim Agreement: No Vendor Name and Address: Fort Defiance Industries LLC 2411 Maremont Pkwy Loudon Tennessee United States 37774 Nature of Requirements: Field Steam Sterilizers W6369-20A005/A Lamorie, Cindy Telephone No. - (613) 687-6655 Fax No. - (613) 687-6656 1. Advance Contract Award Notice (ACAN) An ACAN is a public notice indicating to the supplier community that a department or agency intends to award a contract for goods, services or construction to a pre-identified supplier, thereby allowing other suppliers to signal their interest in bidding, by submitting a statement of capabilities. If no supplier submits a statement of capabilities that meets the requirements set out in the ACAN, on or before the closing date stated in the ACAN, the contracting officer may then proceed with the award to the pre-identified supplier. 2. Definition of the Requirement The Department of National Defence has a requirement for the supply of three (3) fully functional model P2131 Field Sterilizer units or equivalent Field Sterilizer Units with accessories, including an option to purchase up to an additional five (5) units with accessories over the next three (3) years to support both domestic and deployed operational requirements. 3. Criteria for assessment of the Statement of Capabilities (Minimum Essential Requirements) Any interested supplier must demonstrate by way of a statement of capabilities that its Field Sterilizer Unit meets the following requirements: As Canada is a member of the North Atlantic Treaty Organization (NATO) and the Government of Canada is a signatory to NATO Standardization Agreement (STANAG) 2906 AMedP-1.13 - Essential Physical Requirements and Performance Characteristics of Field Type High Pressure Steam Sterilizers, the Fort Defiance Industries model P2131 Field Sterilizer or equivalent must have the following capabilities: M.1 The device must meet the requirements of NATO Standardization Agreement (STANAG) AMedP-1.13: Essential Physical Requirements and Performance Characteristics of Field Type High Pressure Steam Sterilizers. Please note: As NATO is not a certification body the CF H Svcs Gp Technical Authority will determine compliance based on Fort Defiance Industries supplied documentation and provide the Contracting Authority with written assurance the P2131 Field Sterilizer is NATO compliant. M.2 The device must have a valid, active, medical device license issued by Health Canada. Proof of licensure must be included with the bid

submission. The Canadian Armed Forces reserves the right, at its discretion, to verify the validity of the device license through Health Canada's Medical Device Active License Listing website <https://health-products.canada.ca/mdall-limh/index-eng.jsp> M.3 The device must meet or exceed ANSI/AAMI ST8:2013 Hospital Steam Sterilizers <http://my.aami.org/store/?navItemNumber=1338>. GENERAL CHARACTERISTIC The device MUST include: M4 Micro-processor controlled sterilizer cycles M5 Incorporate pre and post vacuum dynamic air removal M6 Alarm functions designed to abort a cycle in the event of abnormal operating conditions M7 Fully automated sterilization cycles, preprogrammed for: a. Immediate use b. Textiles c. Wrapped instruments d. Bowie Dick and vacuum leak test PHYSICAL SPECIFICATIONS Sterilizer Unit: M8 The sterilizer housing must: a. create a self-contained shipping crate; b. include endcaps which protect the chamber door and rear electrical/plumbing connections during transport but capable of folding downward and locking into position to create a stand once in the operating position; and c. include a minimum of four (4) handles, 2 handles per side, to assist in moving. M9 Minimum chamber dimensions: 16 inch diameter by 36 inches in length M10 Maximum overall dimensions (packaged for transport): 46 inches in length by 26.75 inches in width, 31 inches in height with tolerance limit of +/- one (1) inch to the overall dimensions M11 Maximum dimensions (open for operation): 38 inches in length by 26.75 inches in width by 62 inches in height with tolerance limit of +/- one (1) inch to the overall dimensions M12 Sterilization temperature for all cycles, 270 degrees Fahrenheit (132 degrees Celsius) M13 Maximum working pressure of 40 pounds per square inch M14 Jacket water capacity 4.25 US gallons (16 litres) M15 Power requirements: 208-230 volts of alternating current, three (3) phase, 30 amps at 50/60 Hertz M16 Maximum empty weight of 325 pounds M17 Door must include a safety preventing a cycle to start unless the door is closed and latched M18 Water Recovery System: a. Must fit under the sterilizer unit when the sterilizer is setup for operation (the endcaps are open and locked in position) b. Connect to the sterilizer unit to recover water/steam from each cycle c. Provide the means by which the sterilizer unit is able to create pre and post cycle vacuum d. Maximum empty weight of 180 pounds e. Power requirements: 115 volts of alternating current, one (1) phase f. Minimum water storage capacity of 6 US gallons (22 litres) g. Permit the sterilizer to run a minimum of 100 cycles before requiring water to be added M19 Sterilizer Water Softener: a. Remove minerals from the water prior to being added to the Water Recovery System. This will help eliminate scale buildup on the heating elements and within the sterilizer jacket b. Connect to a normal household pressurized potable water source c. Include a hand pump in the event a pressurized water source is not available d. Maximum weight of 35 pounds 4. Applicability of the trade agreement(s) to the procurement This procurement is subject to the following trade agreement(s): o Canadian Free Trade Agreement (CFTA) o North American Free Trade Agreement (NAFTA) o World Trade Organization - Agreement on Government Procurement (WTO-AGP) 5. Justification for the Pre-Identified Supplier The Fort Defiance Industries model P2131 Field Sterilizer unique technical feature is that it is the only self-contained, transportable, Field Sterilizer unit capable of programmable, micro-processor controlled, sterilization cycles offering data logging capability thus ensuring the Government of Canada is in compliance with its NATO obligations. 6. Government Contracts Regulations Exception(s) The following exception(s) to the Government Contracts Regulations is (are) invoked for this procurement under subsection 6(d) - "only one person is capable of performing the work". 7. Exclusions and/or Limited Tendering Reasons The following exclusion(s) and/or limited tendering reasons are invoked under: Canadian Free Trade Agreement (CFTA) Article 513 (b) if the goods or services can be supplied only by a particular supplier and no reasonable alternative or substitute goods or services exist for any of the following reasons: (iii) due to an absence of competition for technical reasons; North American Free Trade Agreement (NAFTA) Article 1016 2. An entity may use limited tendering procedures in the following circumstances and subject to the following conditions, as applicable: (b) where, for works of art, or for reasons connected with the protection of patents, copyrights or other exclusive rights, or proprietary information or where there is an absence of competition for technical reasons, the goods or services can be supplied only by a particular supplier and no reasonable alternative or substitute exists; World Trade Organization Agreement on Government Procurement (WTO-AGP) Article XIII - Limited Tendering (b, iii.) Where the goods or services can be supplied only by a particular supplier and no reasonable alternative or substitute goods or services exist for any of the following reasons: due to an absence of competition for technical reasons; 8. Ownership of Intellectual Property There are not IP terms. 9. Period of the proposed contract or delivery date The equipment must be delivered on or before 30 September 2019. 10. Name and address of the pre-identified supplier: Fort Defiance Industries (FDI) 2411 Maremont Parkway Loudon, TN 37774 11. Suppliers' right to submit a statement of capabilities Suppliers who consider themselves fully qualified and available to provide the goods, services or construction services described in the ACAN may submit a statement of capabilities in writing to the contact person identified in this notice on or before the closing date of this notice. The statement of capabilities must clearly demonstrate how the supplier meets the advertised requirements. 12. Closing date for a submission of a statement of capabilities The closing date and time for accepting statements of capabilities is 23 August 2019, at 2:00 p.m. (EDT). 13. Inquiries and submission of statements of capabilities Inquiries and statements of capabilities are to be directed to: Cynthia Lamorie Supply Specialist Public Works and Government Services Canada Acquisitions Address: Petawawa, Ontario, K8H 2X3 Telephone: 613-401-3643 E-mail address: cynthia.lamorie@pwgsc-tpsgc.gc.ca Delivery Date: Above-mentioned You are hereby notified that the government intends to negotiate with one firm only as identified above. Should you have any questions concerning this requirement, contact the contracting officer identified above. An Advance Contract Award Notice (ACAN) allows departments and agencies to post a notice, for no less than fifteen (15) calendar days, indicating to the supplier community that it intends to award a good, service or construction contract to a pre-identified contractor. If no other supplier submits, on or before the closing date, a Statement of Capabilities that meets the requirements set out in the ACAN, the contracting authority may then proceed with the award. However, should a Statement of Capabilities be found to meet the requirements set out in the ACAN, then the contracting authority will proceed to a full tendering process. Suppliers who consider themselves fully qualified and available to provide the services/goods described

herein, may submit a statement of capabilities in writing to the contact person identified in this Notice on or before the closing date of this Notice. The statement of capabilities must clearly demonstrate how the supplier meets the advertised requirements. The PWGSC file number, the contracting officer's name and the closing date of the ACAN must appear on the outside of the envelope in block letters or, in the case of a facsimile transmission, on the covering page. The Crown retains the right to negotiate with suppliers on any procurement. Documents may be submitted in either official language of Canada.

KOMPASS MARCHÉS PUBLICS INTERNATIONAUX